
Midmark M11 - v1.0.5
Total Cycles: 482

Name: RITTER2
Sterilizer ID: V2313228
10/26/2021 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 74.1 F 0.0
4:00 76.8 F 0.0
6:00 94.9 F 0.1
8:00 140.2 F 0.2
10:00 169.4 F 0.4
12:00 202.5 F 2.3
14:00 211.4 F 5.4
16:00 217.1 F 8.8
18:00 227.3 F 12.7
20:00 229.5 F 9.2
22:00 238.6 F 11.4
24:00 243.1 F 12.6
26:00 242.4 F 11.8
28:00 254.4 F 17.9
30:00 264.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.7
2:00 272.5 F 29.7
3:00 272.4 F 29.7
4:00 271.8 F 29.4

Min 270.9 F 28.6
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 224.4 F 0.0
10:00 191.4 F 0.0
15:00 210.1 F 0.0
20:00 210.8 F 0.0
25:00 161.9 F 0.0
30:00 162.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 31:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:10:07
CYCLE COMPLETE

10/26/2021 09:18 AM
ACCEPTED: X
REJECTED:
Operator: U01