

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1909

Name: RITTER 1  
Sterilizer ID: V2178198  
05/04/2022 12:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.2 F 0.0  
2:00 170.0 F 3.5  
4:00 187.8 F 7.7  
6:00 208.0 F 12.8  
8:00 219.3 F 6.7  
10:00 232.9 F 11.8  
12:00 233.0 F 8.5  
14:00 240.9 F 10.6  
16:00 245.7 F 13.2  
18:00 258.1 F 19.9  
20:00 269.3 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.2  
2:00 271.5 F 28.4  
3:00 272.4 F 29.3  
4:00 272.4 F 28.9  
  
Min 271.4 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.6 F 0.0  
5:00 217.3 F 0.0  
10:00 231.5 F 0.0  
15:00 231.4 F 0.0  
20:00 229.1 F 0.0  
25:00 229.7 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 21:05  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:13  
CYCLE COMPLETE

05/04/2022 01:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_