
Midmark M11 - v1.0.5
Total Cycles: 3510

Name: RITTER2
Sterilizer ID: V2313228
04/09/2024 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.6 F 0.0
2:00 145.2 F 2.0
4:00 166.1 F 5.4
6:00 193.5 F 10.1
8:00 209.7 F 5.2
10:00 224.3 F 8.4
12:00 236.8 F 13.6
14:00 237.8 F 11.0
16:00 237.1 F 9.7
18:00 251.0 F 16.5
20:00 263.0 F 23.4
22:00 272.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.1
1:00 272.2 F 29.9
2:00 272.0 F 29.2
3:00 272.0 F 29.8
4:00 272.5 F 29.5

Min 271.2 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 228.9 F 0.0
10:00 207.2 F 0.0
15:00 205.4 F 0.0
20:00 206.6 F 0.0
25:00 206.5 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

04/09/2024 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____