
Midmark M11 - v1.0.5
Total Cycles: 2280

Name: RITTER2
Sterilizer ID: V2313228
04/21/2023 08:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.5 F 0.0
2:00 81.0 F 0.0
4:00 103.8 F 0.0
6:00 124.4 F 0.0
8:00 148.1 F 2.4
10:00 176.2 F 5.3
12:00 195.9 F 8.5
14:00 209.4 F 12.5
16:00 216.5 F 5.3
18:00 226.9 F 9.3
20:00 237.8 F 12.8
22:00 237.4 F 10.4
24:00 238.0 F 8.8
26:00 247.6 F 14.4
28:00 258.7 F 20.5
30:00 269.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.4 F 29.6
2:00 271.8 F 28.9
3:00 272.2 F 29.5
4:00 272.0 F 29.1

Min 271.3 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 203.9 F 0.0
10:00 197.5 F 0.0
15:00 194.7 F 0.0
20:00 210.6 F 0.0
25:00 217.5 F 0.0
30:00 214.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:56
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:02
CYCLE COMPLETE

04/21/2023 09:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____