

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2448

Name: RITTER 1  
Sterilizer ID: V2178198  
10/18/2022 10:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 220.1 F 0.0  
2:00 213.1 F 4.2  
4:00 211.2 F 10.5  
6:00 224.2 F 6.4  
8:00 239.0 F 12.4  
10:00 238.1 F 10.5  
12:00 237.8 F 9.4  
14:00 252.1 F 16.6  
16:00 264.1 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.3  
2:00 271.5 F 28.5  
3:00 272.8 F 29.4  
4:00 272.1 F 28.8  
  
Min 271.5 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 229.1 F 0.0  
10:00 226.2 F 0.0  
15:00 229.4 F 0.0  
20:00 228.5 F 0.0  
25:00 229.1 F 0.0  
30:00 231.0 F 0.0  
Drying Complete

FILLING: 3:16  
HEATING: 17:58  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:31  
CYCLE COMPLETE

10/18/2022 11:36 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_