
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001613
15:05 11/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:40
PRESSURIZING 6:27
STERILIZING 7:59
136.0 C 220kPa 7:59
136.5 C 221kPa 8:29
136.2 C 221kPa 8:59
136.2 C 220kPa 9:30
136.4 C 222kPa 10:00
136.5 C 223kPa 10:30
136.5 C 223kPa 11:00
136.4 C 222kPa 11:29
VENTING 11:30
AIR DRYING 12:38
CYCLE COMPLETE 27:40

Digital Signature #
0C16D358BD04D7AB
