

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1157

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2022 08:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.3 F 0.0  
2:00 169.4 F 2.1  
4:00 167.4 F 5.7  
6:00 187.0 F 10.4  
8:00 213.1 F 5.1  
10:00 225.9 F 9.7  
12:00 234.3 F 9.2  
14:00 240.8 F 12.4  
16:00 240.2 F 10.9  
18:00 252.4 F 17.3  
20:00 263.4 F 23.9  
22:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.7  
1:00 272.7 F 30.0  
2:00 271.8 F 29.9  
3:00 272.4 F 30.0  
4:00 271.7 F 29.9

Min 271.5 F 29.5  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.4 F 0.0  
5:00 226.3 F 0.0  
10:00 238.9 F 0.0  
15:00 237.2 F 0.0  
20:00 228.7 F 0.0  
25:00 228.0 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 23:01  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:10  
CYCLE COMPLETE

06/06/2022 09:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_