
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 002702
12:33 17/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:36
STERILIZING 3:36
136.0 C 220kPa 3:36
136.1 C 224kPa 4:06
136.4 C 225kPa 4:36
136.4 C 224kPa 5:06
136.4 C 224kPa 5:36
136.3 C 224kPa 6:06
136.4 C 221kPa 6:36
136.0 C 222kPa 7:06
136.0 C 222kPa 7:06
VENTING 7:06
AIR DRYING 8:00
CYCLE COMPLETE 8:04

Digital Signature #
B853397FDA745D5A
