
Midmark M11 - v1.0.0
Total Cycles: 4507

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2024 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 88.1 F 0.0
4:00 110.0 F 0.5
6:00 115.1 F 2.3
8:00 141.2 F 4.9
10:00 165.4 F 8.0
12:00 224.1 F 11.4
14:00 198.6 F 3.3
16:00 228.5 F 7.4
18:00 237.1 F 11.2
20:00 224.1 F 5.8
22:00 241.7 F 11.1
24:00 235.0 F 8.1
26:00 248.8 F 14.4
28:00 260.0 F 20.6
30:00 270.4 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.1 F 28.8
2:00 272.4 F 28.9
3:00 272.5 F 28.8
4:00 272.4 F 28.8

Min 271.7 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 201.1 F 0.0
10:00 208.9 F 0.0
15:00 224.9 F 0.0
20:00 239.1 F 0.0
25:00 228.7 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:21
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:30
CYCLE COMPLETE

05/01/2024 09:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____