
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001241
11:47 01/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:07
STERILIZING 4:17
136.0 C 220kPa 4:17
136.0 C 221kPa 4:47
136.3 C 223kPa 5:17
136.2 C 223kPa 5:47
136.4 C 225kPa 6:17
136.3 C 224kPa 6:47
136.4 C 224kPa 7:17
136.2 C 223kPa 7:47
136.2 C 223kPa 7:47
VENTING 7:47
AIR DRYING 8:39
CYCLE COMPLETE 23:39

Digital Signature #
9286C23AED769F2C
