
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001280
8:32 03/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:55
STERILIZING 5:08
136.0 C 220kPa 5:08
136.4 C 224kPa 5:38
136.3 C 224kPa 6:08
136.4 C 225kPa 6:38
136.3 C 223kPa 7:08
136.3 C 224kPa 7:38
136.2 C 222kPa 8:08
136.4 C 224kPa 8:38
136.4 C 224kPa 8:38
VENTING 8:38
AIR DRYING 9:34
CYCLE COMPLETE 11:43

Digital Signature #
27C04598A0E21C8D
