
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003579
15:23 02/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:27
STERILIZING 3:32
136.4 C 220kPa 3:32
136.4 C 223kPa 4:02
136.4 C 220kPa 4:32
136.4 C 223kPa 5:02
136.4 C 223kPa 5:32
136.1 C 223kPa 6:02
136.4 C 220kPa 6:32
136.0 C 222kPa 7:02
136.0 C 222kPa 7:02
VENTING 7:02
AIR DRYING 7:52
CYCLE COMPLETE 7:55

Digital Signature #
B8F52CB44A057046
