

-----  
Midmark M11 - v1.0.5  
Total Cycles: 686

Name: RITTER2  
Sterilizer ID: V2313228  
01/12/2022 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
2:00 188.9 F 3.2  
4:00 208.8 F 7.7  
6:00 225.0 F 12.9  
8:00 225.3 F 8.1  
10:00 236.2 F 11.9  
12:00 239.0 F 11.8  
14:00 239.2 F 10.8  
16:00 252.9 F 17.7  
18:00 264.8 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.4 F 30.2  
2:00 272.3 F 29.9  
3:00 271.8 F 29.4  
4:00 271.3 F 29.5  
  
Min 271.2 F 29.2  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 218.9 F 0.0  
10:00 232.2 F 0.0  
15:00 227.3 F 0.0  
20:00 221.9 F 0.0  
25:00 220.2 F 0.0  
30:00 219.1 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 19:57  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

01/12/2022 12:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_