

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008031  
8:42 07/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:19  
STERILIZING 4:20  
136.4 C 221kPa 4:20  
136.2 C 221kPa 4:50  
136.3 C 223kPa 5:20  
136.4 C 221kPa 5:50  
136.4 C 222kPa 6:20  
136.1 C 221kPa 6:50  
136.3 C 222kPa 7:20  
136.3 C 221kPa 7:50  
136.3 C 221kPa 7:50  
VENTING 7:50  
AIR DRYING 8:42  
CYCLE COMPLETE 29:08

-----  
Digital Signature #  
4CDDC51A44E4E5A4  
-----