
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001494
11:28 04/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:20
STERILIZING 3:20
136.0 C 220kPa 3:20
136.3 C 224kPa 3:50
136.2 C 223kPa 4:20
136.4 C 224kPa 4:50
136.4 C 225kPa 5:20
136.4 C 225kPa 5:50
135.9 C 221kPa 6:20
136.4 C 224kPa 6:50
136.4 C 224kPa 6:50
VENTING 6:51
AIR DRYING 7:41
CYCLE COMPLETE 37:41

Digital Signature #
992A0B4C461CF540
