
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000963
15:40 07/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:19
PRESSURIZING 1:28
STERILIZING 2:30
136.0 C 227kPa 2:30
135.8 C 229kPa 3:00
136.3 C 231kPa 3:30
136.4 C 232kPa 4:00
136.4 C 231kPa 4:30
136.2 C 231kPa 5:00
136.0 C 229kPa 5:30
136.2 C 231kPa 6:00
136.2 C 231kPa 6:00
VENTING 6:00
AIR DRYING 6:51
CYCLE COMPLETE 21:33

Digital Signature #
ACD519AB5133BFE5
