
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001542
16:06 10/28/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:32
PRESSURIZING 6:17
STERILIZING 7:51
136.0 C 220kPa 7:51
136.4 C 219kPa 8:21
136.4 C 223kPa 8:51
136.3 C 222kPa 9:21
136.5 C 221kPa 9:51
136.5 C 223kPa 10:21
136.4 C 220kPa 10:52
136.3 C 222kPa 11:21
VENTING 11:21
AIR DRYING 12:25
CYCLE COMPLETE 13:06

Digital Signature #
9497B54CBC75C3C8
