

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3059

Name: RITTER2  
Sterilizer ID: V2313228  
10/27/2023 01:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.3 F 0.0  
2:00 132.2 F 1.1  
4:00 158.1 F 4.1  
6:00 180.5 F 8.4  
8:00 202.3 F 13.4  
10:00 219.6 F 7.2  
12:00 231.7 F 12.3  
14:00 233.9 F 9.5  
16:00 236.9 F 9.4  
18:00 247.9 F 14.8  
20:00 259.9 F 21.6  
22:00 270.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.6 F 29.7  
2:00 271.5 F 29.3  
3:00 272.5 F 29.8  
4:00 271.6 F 29.1

Min 271.4 F 29.0  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
5:00 203.7 F 0.0  
10:00 211.4 F 0.0  
15:00 210.1 F 0.0  
20:00 207.8 F 0.0  
25:00 209.1 F 0.0  
30:00 208.4 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 22:44  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:01:15  
CYCLE COMPLETE

10/27/2023 02:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_