
Midmark M11 - v1.0.0
Total Cycles: 3581

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2023 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.6 F 0.0
2:00 161.2 F 0.1
4:00 183.9 F 2.5
6:00 200.4 F 5.9
8:00 213.1 F 9.9
10:00 223.8 F 8.5
12:00 229.7 F 8.5
14:00 240.8 F 13.3
16:00 240.0 F 10.8
18:00 237.5 F 9.2
20:00 251.0 F 15.5
22:00 262.0 F 21.8
24:00 272.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.4 F 28.6
2:00 271.3 F 28.1
3:00 272.6 F 28.7
4:00 271.5 F 28.0

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 207.0 F 0.0
10:00 215.5 F 0.0
15:00 215.2 F 0.0
20:00 215.4 F 0.0
25:00 213.5 F 0.0
30:00 208.9 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 24:30
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:07
CYCLE COMPLETE

07/25/2023 04:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____