
Midmark M11 - v1.0.5
Total Cycles: 3740

Name: RITTER2
Sterilizer ID: V2313228
06/27/2024 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 71.1 F 0.0
4:00 76.9 F 0.1
6:00 114.7 F 0.4
8:00 170.2 F 3.5
10:00 164.8 F 6.9
12:00 195.1 F 10.4
14:00 208.3 F 7.0
16:00 219.6 F 7.2
18:00 229.6 F 11.3
20:00 229.1 F 7.0
22:00 239.9 F 12.3
24:00 239.0 F 10.5
26:00 251.5 F 16.7
28:00 262.3 F 23.0
30:00 272.5 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.5 F 29.7
2:00 271.7 F 29.6
3:00 272.3 F 29.4
4:00 272.2 F 29.9

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 211.0 F 0.0
10:00 192.5 F 0.0
15:00 187.2 F 0.0
20:00 182.2 F 0.0
25:00 180.3 F 0.0
30:00 189.6 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:25
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:08:37
CYCLE COMPLETE

06/27/2024 10:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____