

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4672

Name: RITTER 1  
Sterilizer ID: V2178198  
06/24/2024 01:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.6 F 0.0  
2:00 78.4 F 0.1  
4:00 130.7 F 0.3

C103: HEATUP MODE

\*\* ITEMS NOT STERILE \*\*

06/24/2024 02:03 PM