

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5495

Name:  
Sterilizer ID: V2178198  
04/10/2025 01:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.6 F 0.0  
2:00 114.9 F 1.4  
4:00 144.6 F 4.5  
6:00 165.4 F 8.3  
8:00 188.5 F 12.8  
10:00 209.6 F 4.8  
12:00 220.3 F 8.8  
14:00 232.6 F 13.1  
16:00 233.7 F 8.7  
18:00 243.7 F 12.2  
20:00 242.7 F 11.7  
22:00 254.7 F 17.8  
24:00 265.4 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 29.0  
2:00 271.3 F 28.2  
3:00 272.4 F 29.1  
4:00 272.0 F 28.4

Min 271.3 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 200.6 F 0.0  
10:00 192.7 F 0.0  
15:00 189.9 F 0.0  
20:00 189.3 F 0.0  
25:00 213.0 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:48  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:04:16  
CYCLE COMPLETE

04/10/2025 02:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_