
Midmark M11 - v1.0.0
Total Cycles: 1248

Name: RITTER 1
Sterilizer ID: V2178198
10/08/2021 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.7 F 0.0
2:00 180.5 F 2.5
4:00 196.3 F 6.8
6:00 227.2 F 12.8
8:00 228.7 F 7.7
10:00 240.3 F 13.2
12:00 238.9 F 10.5
14:00 237.8 F 8.6
16:00 251.0 F 15.7
18:00 263.1 F 22.7
20:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.7 F 28.3
3:00 272.6 F 28.8
4:00 271.6 F 28.1

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.1 F 0.0
5:00 205.2 F 0.0
10:00 201.4 F 0.0
15:00 195.2 F 0.0
20:00 186.7 F 0.0
25:00 181.5 F 0.0
30:00 185.2 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

10/08/2021 10:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____