
Midmark M11 - v1.0.5
Total Cycles: 2840

Name: RITTER2
Sterilizer ID: V2313228
09/06/2023 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.5 F 0.0
2:00 177.8 F 3.7
4:00 200.6 F 9.0
6:00 221.2 F 7.8
8:00 231.7 F 10.4
10:00 232.8 F 8.0
12:00 240.4 F 10.8
14:00 247.4 F 14.4
16:00 260.2 F 21.5
18:00 271.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.4 F 29.3
2:00 272.2 F 29.6
3:00 271.9 F 29.0
4:00 272.4 F 29.7

Min 271.3 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 233.4 F 0.0
10:00 230.2 F 0.0
15:00 224.6 F 0.0
20:00 222.1 F 0.0
25:00 221.2 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 18:34
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

09/06/2023 04:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____