
Midmark M11 - v1.0.0
Total Cycles: 4451

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2024 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.1 F 0.0
2:00 161.8 F 1.6
4:00 164.1 F 4.7
6:00 182.6 F 8.3
8:00 200.9 F 12.2
10:00 213.0 F 4.5
12:00 221.6 F 8.4
14:00 232.5 F 12.5
16:00 233.8 F 9.1
18:00 240.3 F 11.0
20:00 244.3 F 12.9
22:00 255.7 F 18.8
24:00 265.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.5
2:00 271.9 F 29.0
3:00 271.5 F 29.1
4:00 272.4 F 29.5

Min 271.2 F 28.7
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 206.1 F 0.0
10:00 207.0 F 0.0
15:00 207.2 F 0.0
20:00 207.8 F 0.0
25:00 211.3 F 0.0
30:00 213.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:49
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:38
CYCLE COMPLETE

04/12/2024 10:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____