
Midmark M11 - v1.0.0
Total Cycles: 2722

Name: RITTER 1
Sterilizer ID: V2178198
12/28/2022 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.9 F 0.0
2:00 79.3 F 0.0
4:00 85.0 F 0.0
6:00 112.0 F 0.8
8:00 144.1 F 2.4
10:00 172.9 F 4.5
12:00 192.1 F 7.7
14:00 206.3 F 11.2
16:00 217.6 F 4.7
18:00 224.4 F 7.8
20:00 236.9 F 12.0
22:00 234.3 F 8.6
24:00 241.5 F 11.0
26:00 244.7 F 12.7
28:00 256.5 F 18.9
30:00 267.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.2
2:00 271.7 F 28.4
3:00 272.6 F 29.3
4:00 272.1 F 28.7

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 218.6 F 0.0
10:00 221.3 F 0.0
15:00 220.8 F 0.0
20:00 228.4 F 0.0
25:00 229.8 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:31
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

12/28/2022 09:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____