
Midmark M11 - v1.0.5
Total Cycles: 1863

Name: RITTER2
Sterilizer ID: V2313228
12/20/2022 12:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.8 F 0.0
2:00 147.1 F 0.0
4:00 159.9 F 0.0
6:00 200.1 F 1.0
8:00 211.7 F 5.9
10:00 225.9 F 10.9
12:00 229.7 F 7.2
14:00 243.4 F 13.3
16:00 243.1 F 12.1
18:00 243.0 F 11.7
20:00 256.8 F 19.0
22:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.2
2:00 271.7 F 28.3
3:00 272.3 F 29.2
4:00 272.4 F 28.9

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 213.9 F 0.0
10:00 231.4 F 0.0
15:00 226.6 F 0.0
20:00 227.0 F 0.0
25:00 230.8 F 0.0
30:00 231.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:08
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

12/20/2022 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____