
Midmark M11 - v1.0.0
Total Cycles: 3763

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2023 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.3 F 0.0
2:00 188.2 F 4.0
4:00 206.1 F 9.8
6:00 221.3 F 5.9
8:00 235.3 F 12.4
10:00 237.5 F 10.1
12:00 235.5 F 8.6
14:00 251.7 F 16.2
16:00 264.5 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 271.9 F 28.3
2:00 272.5 F 28.8
3:00 272.1 F 28.8
4:00 271.8 F 28.3

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 223.8 F 0.0
10:00 224.7 F 0.0
15:00 235.3 F 0.0
20:00 237.3 F 0.0
25:00 228.2 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 17:45
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:17
CYCLE COMPLETE

09/12/2023 06:00 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____