
Midmark M11 - v1.0.5
Total Cycles: 1303

Name: RITTER2
Sterilizer ID: V2313228
07/15/2022 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.8 F 0.0
2:00 150.1 F 0.0
4:00 153.8 F 0.3
6:00 190.7 F 3.7
8:00 207.1 F 8.3
10:00 218.1 F 13.3
12:00 226.6 F 8.5
14:00 238.4 F 13.5
16:00 238.6 F 11.3
18:00 237.7 F 9.8
20:00 250.6 F 16.3
22:00 261.9 F 22.8
24:00 272.3 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.5 F 29.6
2:00 271.6 F 29.5
3:00 272.4 F 29.4
4:00 271.7 F 29.6

Min 271.4 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 208.6 F 0.0
10:00 218.9 F 0.0
15:00 221.0 F 0.0
20:00 219.0 F 0.0
25:00 217.8 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

07/15/2022 09:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____