
Midmark M11 - v1.0.5
Total Cycles: 1736

Name: RITTER2
Sterilizer ID: V2313228
11/16/2022 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.0 F 0.0
2:00 191.7 F 1.9
4:00 218.8 F 7.3
6:00 233.5 F 13.0
8:00 236.5 F 9.8
10:00 236.3 F 8.6
12:00 241.8 F 11.0
14:00 248.1 F 14.5
16:00 261.0 F 21.8
18:00 272.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.3 F 29.3
2:00 271.6 F 28.9
3:00 272.3 F 29.4
4:00 271.6 F 28.7

Min 271.5 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 227.9 F 0.0
10:00 247.1 F 0.0
15:00 242.4 F 0.0
20:00 231.6 F 0.0
25:00 226.3 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:26
CYCLE COMPLETE

11/16/2022 04:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____