
Midmark M11 - v1.0.0
Total Cycles: 3329

Name: RITTER 1
Sterilizer ID: V2178198
06/02/2023 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.3 F 0.0
2:00 132.1 F 0.0
4:00 146.4 F 0.4
6:00 174.1 F 3.7
8:00 200.5 F 7.1
10:00 212.3 F 10.8
12:00 222.6 F 6.2
14:00 232.2 F 9.6
16:00 237.6 F 10.2
18:00 239.4 F 11.1
20:00 237.4 F 8.9
22:00 249.3 F 15.0
24:00 260.1 F 20.9
26:00 270.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 29.0
2:00 271.3 F 28.3
3:00 272.2 F 29.1
4:00 272.2 F 28.8

Min 271.0 F 28.2
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 204.9 F 0.0
10:00 226.0 F 0.0
15:00 230.0 F 0.0
20:00 228.2 F 0.0
25:00 226.8 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:55
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:05:27
CYCLE COMPLETE

06/02/2023 03:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____