
Midmark M11 - v1.0.0
Total Cycles: 2526

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2022 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.9 F 0.0
2:00 88.1 F 0.0
4:00 93.9 F 0.0
6:00 117.9 F 0.1
8:00 152.6 F 2.8
10:00 178.8 F 6.3
12:00 195.5 F 10.1
14:00 213.3 F 7.5
16:00 221.3 F 6.6
18:00 231.7 F 10.5
20:00 234.4 F 7.8
22:00 242.0 F 11.7
24:00 238.9 F 9.5
26:00 251.4 F 15.6
28:00 262.0 F 22.0
30:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.6 F 28.7
2:00 272.1 F 28.6
3:00 272.3 F 28.5
4:00 272.2 F 28.7

Min 271.6 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 212.2 F 0.0
10:00 232.6 F 0.0
15:00 219.3 F 0.0
20:00 221.2 F 0.0
25:00 219.5 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:23
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:15
CYCLE COMPLETE

11/08/2022 09:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____