

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.8uS / 0.5ppm  
CYCLE NUMBER 001434  
12:00 09/15/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:31  
PRESSURIZING 6:29  
STERILIZING 8:22  
136.0 C 220kPa 8:22  
136.5 C 222kPa 8:52  
136.5 C 223kPa 9:22  
136.5 C 221kPa 9:52  
136.4 C 222kPa 10:22  
136.4 C 221kPa 10:52  
136.2 C 221kPa 11:23  
136.1 C 221kPa 11:52  
VENTING 11:52  
AIR DRYING 12:58  
CYCLE COMPLETE 36:27  
-----

Digital Signature #  
D183D75963859F65  
-----