

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000317  
9:00 04/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:29  
PRESSURIZING 3:36  
STERILIZING 4:55  
136.0 C 229kPa 4:55  
136.4 C 232kPa 5:25  
136.1 C 230kPa 5:55  
136.4 C 230kPa 6:25  
136.4 C 230kPa 6:55  
136.1 C 228kPa 7:25  
136.1 C 229kPa 7:55  
136.2 C 230kPa 8:25  
136.2 C 230kPa 8:25  
VENTING 8:25  
AIR DRYING 9:18  
CYCLE COMPLETE 22:32

-----  
Digital Signature #  
EF69143A848A37EE  
-----