
Midmark M11 - v1.0.0
Total Cycles: 920

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2021 09:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 193.6 F 2.5
4:00 201.3 F 6.5
6:00 214.8 F 11.8
8:00 222.7 F 6.3
10:00 235.7 F 11.4
12:00 234.9 F 8.7
14:00 240.2 F 10.5
16:00 246.2 F 13.5
18:00 258.2 F 19.9
20:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.3 F 29.2
2:00 271.9 F 28.7
3:00 272.0 F 29.0
4:00 272.3 F 29.0

Min 271.4 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 224.3 F 0.0
10:00 230.8 F 0.0
15:00 227.0 F 0.0
20:00 225.2 F 0.0
25:00 224.0 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

06/21/2021 11:10 AM
ACCEPTED: X
REJECTED:
Operator: U01