
Midmark M11 - v1.0.0
Total Cycles: 2326

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2022 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.0 F 0.0
2:00 192.3 F 4.4
4:00 210.5 F 9.4
6:00 219.4 F 6.0
8:00 231.6 F 9.8
10:00 237.1 F 8.3
12:00 244.1 F 13.2
14:00 243.3 F 12.1
16:00 256.6 F 19.0
18:00 268.3 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 29.0
2:00 271.8 F 28.8
3:00 272.3 F 29.0
4:00 271.6 F 28.6

Min 271.6 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 229.1 F 0.0
10:00 234.1 F 0.0
15:00 230.9 F 0.0
20:00 225.8 F 0.0
25:00 223.5 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

09/14/2022 02:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____