
Midmark M11 - v1.0.5
Total Cycles: 2957

Name: RITTER2
Sterilizer ID: V2313228
10/02/2023 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.2 F 0.0
2:00 190.2 F 3.8
4:00 202.8 F 9.4
6:00 219.1 F 6.6
8:00 233.1 F 10.4
10:00 234.1 F 8.0
12:00 240.9 F 10.5
14:00 248.7 F 14.7
16:00 262.1 F 22.2
18:00 272.1 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.1 F 28.7
2:00 272.7 F 29.2
3:00 271.5 F 28.6
4:00 272.4 F 28.9

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 214.0 F 0.0
10:00 228.5 F 0.0
15:00 229.4 F 0.0
20:00 225.8 F 0.0
25:00 222.8 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:09
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

10/02/2023 12:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____