

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003610  
9:48 05/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:38  
STERILIZING 3:34  
136.0 C 220kPa 3:34  
136.2 C 221kPa 4:04  
136.2 C 222kPa 4:34  
136.2 C 220kPa 5:04  
136.3 C 223kPa 5:34  
136.4 C 223kPa 6:04  
136.4 C 223kPa 6:34  
136.3 C 223kPa 7:04  
136.3 C 223kPa 7:04  
VENTING 7:04  
AIR DRYING 7:55  
CYCLE COMPLETE 8:10  
-----

Digital Signature #  
4E16CA8CF7AE4B81  
-----