
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001521
10:07 10/21/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:02
PRESSURIZING 5:34
STERILIZING 7:01
136.0 C 220kPa 7:01
136.4 C 223kPa 7:31
136.0 C 219kPa 8:01
136.3 C 221kPa 8:31
136.0 C 221kPa 9:01
136.3 C 218kPa 9:31
136.4 C 223kPa 10:01
136.4 C 218kPa 10:31
VENTING 10:31
AIR DRYING 11:33
CYCLE COMPLETE 41:34

Digital Signature #
FBBE663AD0FEC8A6
