
Midmark M11 - v1.0.5
Total Cycles: 894

Name: RITTER2
Sterilizer ID: V2313228
03/23/2022 07:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 75.0 F 0.0
4:00 103.2 F 0.0
6:00 130.5 F 0.0
8:00 158.3 F 0.2
10:00 188.4 F 0.5
12:00 208.1 F 2.8
14:00 220.5 F 6.2
16:00 231.2 F 9.9
18:00 237.0 F 10.5
20:00 239.7 F 11.3
22:00 237.1 F 9.3
24:00 238.9 F 9.4
26:00 249.8 F 15.2
28:00 261.8 F 22.0
30:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 271.9 F 28.5
2:00 272.3 F 28.8
3:00 272.6 F 29.2
4:00 272.0 F 29.0

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 225.0 F 0.0
10:00 225.3 F 0.0
15:00 222.4 F 0.0
20:00 221.4 F 0.0
25:00 229.3 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 30:20
STERILIZING: 4:00
VENTING: 2:52
DRYING: 30:00
TOTAL CYCLE: 01:08:10
CYCLE COMPLETE

03/23/2022 08:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____