
Midmark M11 - v1.0.0
Total Cycles: 2716

Name: RITTER 1
Sterilizer ID: V2178198
12/22/2022 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.1 F 0.0
2:00 130.1 F 0.0
4:00 162.7 F 0.2
6:00 186.4 F 3.1
8:00 202.4 F 7.1
10:00 220.5 F 11.5
12:00 228.6 F 5.6
14:00 234.4 F 10.4
16:00 237.9 F 7.4
18:00 243.3 F 12.7
20:00 241.4 F 10.9
22:00 254.1 F 17.5
24:00 265.3 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.3
2:00 272.1 F 28.7
3:00 271.3 F 28.6
4:00 272.5 F 29.4

Min 271.2 F 28.3
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 206.3 F 0.0
10:00 219.8 F 0.0
15:00 225.9 F 0.0
20:00 234.1 F 0.0
25:00 227.5 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:48
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

12/22/2022 10:42 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____