
Midmark M11 - v1.0.0
Total Cycles: 4707

Name: RITTER 1
Sterilizer ID: V2178198
07/09/2024 12:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.1 F 0.0
2:00 82.1 F 0.4
4:00 105.5 F 1.8
6:00 140.5 F 4.5
8:00 167.5 F 7.9
10:00 183.9 F 12.0
12:00 203.2 F 3.3
14:00 214.8 F 6.8
16:00 226.2 F 10.6
18:00 229.3 F 7.1
20:00 239.1 F 11.1
22:00 235.5 F 8.5
24:00 249.5 F 15.0
26:00 261.0 F 21.5
28:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.7
2:00 272.2 F 28.9
3:00 272.0 F 28.4
4:00 272.7 F 29.1

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 188.3 F 0.0
10:00 189.6 F 0.0
15:00 197.8 F 0.0
20:00 199.6 F 0.0
25:00 211.2 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:36
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:06:41
CYCLE COMPLETE

07/09/2024 01:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____