
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007520
7:49 10/06/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:08
STERILIZING 3:58
136.0 C 220kPa 3:58
136.2 C 221kPa 4:28
136.2 C 223kPa 4:58
136.4 C 224kPa 5:28
136.2 C 219kPa 5:58
136.1 C 219kPa 6:28
135.9 C 219kPa 6:58
136.3 C 220kPa 7:28
136.3 C 220kPa 7:28
VENTING 7:28
AIR DRYING 8:19
CYCLE COMPLETE 13:02

Digital Signature #
CC8C916FE6DBA973
