
Midmark M11 - v1.0.5
Total Cycles: 1425

Name: RITTER2
Sterilizer ID: V2313228
08/29/2022 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.9 F 0.0
2:00 201.5 F 3.1
4:00 201.4 F 8.4
6:00 219.1 F 10.8
8:00 231.1 F 10.1
10:00 235.0 F 8.5
12:00 242.5 F 11.8
14:00 245.3 F 13.3
16:00 257.4 F 19.9
18:00 268.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.3 F 29.7
2:00 271.2 F 28.9
3:00 272.1 F 29.8
4:00 272.6 F 29.9

Min 271.0 F 28.9
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 231.4 F 0.0
10:00 230.9 F 0.0
15:00 228.8 F 0.0
20:00 227.9 F 0.0
25:00 229.4 F 0.0
30:00 228.7 F 0.0

Drying Complete

FILLING: 1:56
HEATING: 19:15
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

08/29/2022 11:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____