

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5310

Name: RITTER 1  
Sterilizer ID: V2178198  
02/03/2025 02:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.2 F 0.0  
2:00 174.4 F 3.0  
4:00 183.8 F 7.0  
6:00 207.2 F 11.7  
8:00 217.5 F 5.4  
10:00 229.5 F 10.0  
12:00 234.8 F 8.7  
14:00 239.7 F 11.7  
16:00 238.2 F 9.7  
18:00 251.5 F 16.1  
20:00 262.7 F 22.7  
22:00 272.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.6  
1:00 272.3 F 29.3  
2:00 271.9 F 28.7  
3:00 271.9 F 29.1  
4:00 272.3 F 29.1

Min 271.4 F 28.4  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 207.7 F 0.0  
10:00 221.7 F 0.0  
15:00 230.0 F 0.0  
20:00 228.1 F 0.0  
25:00 219.4 F 0.0  
30:00 219.2 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 22:11  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:01:20  
CYCLE COMPLETE

02/03/2025 03:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_