
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001144
16:08 23/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:56
PRESSURIZING 3:59
STERILIZING 5:28
136.0 C 220kPa 5:28
136.4 C 225kPa 5:58
136.2 C 224kPa 6:28
136.1 C 224kPa 6:58
136.3 C 225kPa 7:28
136.1 C 224kPa 7:58
136.1 C 225kPa 8:28
136.4 C 226kPa 8:58
136.4 C 226kPa 8:58
VENTING 8:58
AIR DRYING 9:55
CYCLE COMPLETE 18:42

Digital Signature #
E82C3D4AAC92BBCB
