

-----  
Midmark M11 - v1.0.0  
Total Cycles: 578

Name: MIDMARK  
Sterilizer ID: V2178198  
12/21/2020 12:52 PM

BEGIN PACKS CYCLE  
Temp: 250 Degrees F  
Time: 30 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.9 F 0.0  
2:00 143.9 F 1.2  
4:00 169.7 F 4.0  
6:00 196.4 F 3.8  
8:00 217.6 F 3.9  
10:00 225.0 F 4.7  
12:00 229.5 F 6.6  
14:00 228.4 F 5.5  
16:00 241.6 F 11.1  
18:00 252.7 F 16.4

STERILIZING  
mm:ss Degrees PSI  
0:00 251.1 F 15.9  
1:00 252.4 F 16.4  
2:00 251.1 F 15.9  
3:00 252.6 F 16.5  
4:00 251.8 F 16.1  
5:00 251.3 F 16.1  
6:00 252.4 F 16.5  
7:00 252.3 F 16.3  
8:00 251.5 F 16.0  
9:00 251.3 F 16.0  
10:00 252.2 F 16.6  
11:00 252.5 F 16.5  
12:00 252.4 F 16.3  
13:00 251.9 F 16.2  
14:00 252.1 F 16.4  
15:00 251.7 F 16.2  
16:00 251.7 F 16.3  
17:00 252.2 F 16.7  
18:00 252.6 F 16.9  
19:00 253.2 F 16.9  
20:00 253.2 F 16.8  
21:00 253.2 F 16.7  
22:00 252.9 F 16.6  
23:00 252.6 F 16.5  
24:00 252.2 F 16.4  
25:00 252.0 F 16.3  
26:00 251.7 F 16.2  
27:00 251.6 F 16.2  
28:00 251.9 F 16.4  
29:00 252.4 F 16.7  
30:00 252.9 F 16.9

Min 251.0 F 15.8  
Max 253.3 F 16.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 213.5 F 0.0  
10:00 203.4 F 0.0  
15:00 204.8 F 0.0  
20:00 200.8 F 0.0  
25:00 200.9 F 0.0  
Dry Cycle Aborted

FILLING: 0:58  
HEATING: 18:22

STERILIZING: 30:00  
VENTING: 2:55  
DRYING: 26:25  
TOTAL CYCLE: 01:18:40

12/21/2020 02:11 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01