
Midmark M11 - v1.0.5
Total Cycles: 32

Name: RITTER2
Sterilizer ID: V2313228
05/19/2021 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.2 F 0.0
2:00 174.5 F 3.8
4:00 192.2 F 8.3
6:00 213.0 F 11.0
8:00 226.9 F 9.0
10:00 234.7 F 9.5
12:00 243.4 F 13.2
14:00 244.0 F 12.5
16:00 256.6 F 19.3
18:00 267.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.4 F 29.3
2:00 271.7 F 29.3
3:00 272.4 F 29.1
4:00 271.5 F 29.2

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 231.8 F 0.0
10:00 224.3 F 0.0
15:00 221.7 F 0.0
20:00 219.1 F 0.0
25:00 215.8 F 0.0
30:00 206.6 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

05/19/2021 01:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____