

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1213

Name: RITTER2  
Sterilizer ID: V2313228  
06/22/2022 12:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.4 F 0.0  
2:00 158.0 F 2.9  
4:00 173.6 F 6.7  
6:00 194.9 F 11.3  
8:00 215.7 F 5.5  
10:00 227.8 F 10.8  
12:00 230.7 F 7.7  
14:00 243.0 F 12.6  
16:00 243.6 F 12.6  
18:00 255.9 F 19.1  
20:00 266.9 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.5 F 29.9  
2:00 272.0 F 29.4  
3:00 271.2 F 29.2  
4:00 272.2 F 30.0  
  
Min 271.2 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
5:00 212.3 F 0.0  
10:00 216.3 F 0.0  
15:00 223.3 F 0.0  
20:00 216.2 F 0.0  
25:00 213.7 F 0.0  
30:00 212.1 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 21:35  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:07  
CYCLE COMPLETE

06/22/2022 01:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_