
Midmark M11 - v1.0.5
Total Cycles: 1821

Name: RITTER2
Sterilizer ID: V2313228
12/07/2022 05:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.7 F 0.0
2:00 185.7 F 4.1
4:00 205.1 F 9.7
6:00 221.1 F 6.3
8:00 233.2 F 11.4
10:00 236.3 F 9.7
12:00 237.2 F 9.0
14:00 251.8 F 16.5
16:00 264.4 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.6 F 29.5
2:00 271.9 F 29.3
3:00 272.2 F 29.2
4:00 272.3 F 29.7

Min 271.6 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 233.9 F 0.0
10:00 226.6 F 0.0
15:00 226.2 F 0.0
20:00 227.3 F 0.0
25:00 229.5 F 0.0
30:00 228.6 F 0.0

Drying Complete

FILLING: 2:18
HEATING: 17:40
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:09
CYCLE COMPLETE

12/07/2022 06:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____